



## Declaration of Conformity

According to EN 45014 and Article 10 of Council Directives 2004/108/EC and 73/23/EEC.

We, **MAS Advanced Measurement Solutions** declare under our responsibility that the product,

Weight Indicator  
Model Name: MAS 640 SS+

manufactured at:

MAS Advanced Measurement Solutions  
10 Ballycasey, Court Mews, Shannon  
Co Clare,  
Ireland

is in conformity with the essential requirements of Council Directive 2004/108/EC as regards Electromagnetic compatibility and Council Directive 73/23/EEC as regards Low Voltage Safety based upon compliance of the product with the following harmonised standards:

<b>EMC Standards:</b>			
<b>Emission Standards:</b> PN-EN 6000-6-3:2007 Residential, commercial and light-industrial environments.		<b>Immunity Standards:</b> PN-EN 6000-6-1:2007 Residential, commercial and light-industrial environments.	
Conducted emission measurement	PN-EN 61000-6-3:2007	ESD immunity	PN-EN 61000-4-2:1999+A2:2003
Radiated disturbance measurement	PN-EN 61000-6-3:2007	RF electromagnetic field immunity	PN-EN 61000-4-3:2007+A1:2008
		Bursts	PN-EN 61000-4-4:2005
		Surge 1.2/50µs, 8/20µs	PN-EN 61000-4-5:2006
		RF conducted disturbances	PN-EN 61000-4-6:2007
		Voltage dips and interruption	PN-EN 61000-4-11:2007
<b>Electrical Product Safety Standards:</b>			
PN-EN 60950-1:2004			

Arkadiusz Zurakowski  
Director of Engineering

Date: March 7, 2011